## GS30-PURE II Protocol A \*\* PLEASE DISCARD RED VENTED CAP FROM CONCENTRATING DEVICE BEFORE USE \*\*

Note: Always swab self-sealing port with sterile alcohol prior to accessing with a sterile syringe

